

Certificate of Compliance

We hereby declare that the technical file of product complied with the requirement of directive (2014/34/EU) ATEX Directive.

Certificate No.: UQ-2025093020

Manufacturer

Name : MICROLIT

Address : 629, PAKRAMAU, KURSI ROAD, LUCKNOW, U.P.-226026, INDIA

Products : LIQUID FLUID EQUIPMENT, PERISTALTIC PUMP

Model No : MP110 B, MP310B, MP 610B, MP100B, MP600B, MP300B, MP110C, MP310C, MP610C, MP100C, MP300C, MP600C, MP1101, MP3101, MP6101, MP1001, MP3001, MP6001 VACUUM ASPIRATOR: MODEL NO. V A-1, V A-2 AND V A-4

Applicable Standard : IEC 61010-1, IEC 61326-1, EN 61000-3-2:2019, EN 61000-3-3:2013+A1:2019

The Certification body has performed an audit of the above product quality system covering the design, manufacture and final inspection of the certified product. The quality system has been assessed, approved and is subject to continuous surveillance according to Directive (2014/34/EU) ATEX.

This certificate is issued under the following conditions:

1. It applies only to the quality system maintained in the manufacture of above referenced models and it does not substitute the design or type-examination procedures, if requested.
2. The certificate remains valid until the manufacturing conditions or the quality systems are changed.
3. The certificate validity is conditioned by positive results or surveillance audits.
4. 4. After fulfilling the relevant EU legislation, the manufacturer shall affix to each device, of the above referenced models
5. Compliance with the essential Health and safety Requirement has been assured by compliance with EN 60079 -0:2012/A11:2013, EN60079-1:2017& IEC 60079-1

Validity of this certificate can be verified at www.ukcertifications.org.uk/verify

Date of this certificate

30th September 2025

1st Surveillance Audit Due

29th September 2026

2nd Surveillance Audit Due

29th September 2027

Certificate Expiry (subject to the company maintaining its system to the required standard)

29th September 2028



Authorised Signatory

